

# XD-9502

<https://www.gigahertz-optik.com/zh-cn/product/xd-9502/>

产品标签: UV/紫外线, VIS/愿景,



## 描述

The term **photostability** refers to the ability of a substance or material to remain unchanged under exposure to light. Light as a form of electromagnetic radiation is capable of affecting chemical changes in some substances over time. The pharmaceutical industry must take special precautions to ensure that their drug products and packaging are not adversely affected by light exposure.

"ICH" the International Conference on Harmonization for the pharmaceutical industry publishes guidelines for testing photostability of drugs and drug products.

There are two options for exposing samples to UV, primarily UVA, and white light. Measuring and monitoring the UV irradiance and light levels is necessary to ensure stated exposure (dose) levels for both UVA and light are achieved.

### Exposing Option 1

Any light source that is designed to produce an output similar to the D65/ID65 emission standard such as an artificial daylight fluorescent lamp combining visible and ultraviolet (UV) outputs, xenon, or metal halide lamp. D65 is the internationally recognized standard for outdoor daylight as defined in [ISO 10977](#) (1993). ID65 is the equivalent indoor indirect daylight standard. For a light source emitting significant radiation below 320 nm (nanometers), an appropriate filter(s) may be fitted to eliminate such radiation.

### Exposing Option 2

For option 2 the same sample should be exposed to both the cool white fluorescent and near ultraviolet lamp.


1. A cool white fluorescent lamp designed to produce an output similar to that specified in ISO 18099:2022 (formerly [ISO 10977](#) :1993); and
2. A near UV fluorescent lamp having a spectral distribution from 320 nm to 400 nm with a maximum energy emission between 350 nm and 370 nm; a significant proportion of UV should be in both bands of 320 to 360 nm and 360 to 400 nm.

Also with the UVA sensor flat spectral response and the high fidelity to CIE  $V(\lambda)$  function lux sensor both Option 1 and 2 lamp types can be measured with the same unit without changing calibration factors. For future considerations this would also apply to LED based white light and UV-A sources.

Both detectors have a cosine corrected field of view for accurate spatial responsivity.



UV-A Irradiance and Illuminance Detector



XD-9502-4 UV-A Sensor and  $V(\lambda)$  Sensor - Typical Spectral Responsivities

---

## XD-9502-4 Photostability Light and UV Meter

The light sensor recommended for photostability light exposure testing is the XD-9502-4 dual cell low profile detector that houses both UV-A and lux detectors so both measurements can be made simultaneously. This means no fumbling, no repositioning of the detector for faster measurements and mapping of the exposure plane.

Also with the UVA sensor flat spectral response and the high fidelity to CIE  $v$ -lambda function lux sensor both Option 1 and 2 lamp types can be measured with the same unit without changing calibration factors. The detector shows a cosine-corrected field of view.

**For Use With:** [X1-1 Four-Channel USB Radiometer/Photometer](#)

### Internationally Traceable Calibration

All filter detectors are calibrated to internationally traceable standards following ISO 17025 guidelines and requirements. [Gigahertz-Optik optical radiation calibration laboratory is DIN EN ISO/IEC 17025 accredited](#) (DaKKs Registration No. D-K-15047-01-00).

### Application Case: Here's the problem

You are responsible for toxicity testing of your drug products and drug packaging under light exposure according to ICH Q1B **Photostability Testing of New Active Substances and Medicinal Products**. ICH states that light testing should be an integral part of stress testing of pharmaceuticals and its packaging. The intrinsic photostability characteristics of new drug substances and products should be evaluated to demonstrate that, as appropriate, light exposure does not result in unacceptable change.

### Here are the requirements

According to the guidelines two types of light sources as already described as options 1 and 2 can be used for photostability testing. A pharmaceutical manufacturer/applicant may rely on the spectral distribution specification of the light source manufacturer. A simple and accurate means of measuring UVA and visible light inside the photostability exposure chamber is required for both lamp types.

### Here's the solution

The **X1-1 + XD-9502-4 Photostability Meter** can be used for both lamp types for Options 1 & 2 of the ICH Q1B guidelines.

## 规格

## 规格

光谱响应性	CH1: UV-A (315 - 400) nm CH2: Photopic; $f_1 \leq 6\%$ (380 - 780) nm
输入光学元件	8 mm Ø Cosine diffuser
最大信号电流	CH1: 100 $\mu$ A CH2: 1 mA
典型响应性	CH1: radiometric: 10 nA/(W/m <sup>2</sup> ) CH2: photometric: 50 pA/lx
测量范围	Sensor XD-9502-CH1 min (0.1 pA NEI) 1.67E-05 max (100 $\mu$ A) 1.67E+04 Sensor XD-9502-CH2 min (0.1 pA NEI) 2.00E-03 max (200 $\mu$ A) 4.00E+06
杂项	
温度范围	(5 - 40) °C
电缆长度	2 m

## 采购信息

文章编号	モデル	描述
产品		
15295882	XD-9502 (-4 Connector)	Detector, Calibration Certificate
校准		
15311563	KP-XD9502X1-E-I	Option: DIN EN ISO/IEC 17025:2018 Test Certificate (DAkks). Integral irradiance in the wavelength range 315nm - 400nm and photopic illuminance. In combination with X1 optometer.
重新标定		
15300121	K-XD9502-I	Recalibration with Calibration Certificate
15300416	K-UV-D-SR	Calibration of relative spectral responsivity from 250 nm to 550 nm
15311562	KKP-XD9502X1-E-I	Factory Calibration Certificate with DIN EN ISO/IEC 17025:2018 Test Certificate. In combination with X1 optometer.

## 联系、校准、服务和支持/Contact, Calibration, Service & Support

我们以出色的技术咨询和售后支持而闻名于世。请与我们联系，共同寻找适合您的最佳解决方案。我们的服务。

- 技术咨询和销售
- 售后支持
- 校准和重新校准（[ISO/IEC 17025校准服务](#)，工厂校准，[第三方产品的校准](#)）。
- 维修和更新
- 定制解决方案的OEM和可行性咨询

[请将您的询问发送给我们](#)

(\*为必填项) 或通过电话或电子邮件与我们联系。我们也欢迎你的反馈或在谷歌上评论我们。 [Google](#)

### Gigahertz Optik GmbH (总部)

Tel.: +49 (0)8193-93700-0

Fax: +49 (0)8193-93700-50

[info@gigahertz-optik.de](mailto:info@gigahertz-optik.de)

An der Kaelberweide 12

82299 Tuerkenfeld, Germany